

REPORT DOCUMENTATION PAGE

AFRL-SR-AR-TR-03-

g the
ading
02-
rently

data needed, and completing and reviewing this collection of information. Send comments regarding this burden estimate or any other aspect of this burden to Department of Defense, Washington Headquarters Services, Directorate for Information Operations and Reports (0704-0102). Respondents should be aware that notwithstanding any other provision of law, no person shall be subject to any penalty for failing to comply with a collection of information if it does not have a valid OMB control number. PLEASE DO NOT RETURN YOUR FORM TO THE ABOVE ADDRESS.

0289

1. REPORT DATE (DD-MM-YYYY) 01-07-2003		2. REPORT TYPE Final Technical		3. DATES COVERED (From - To) 01062002 - 31052003	
4. TITLE AND SUBTITLE (U) (DURIP02) Portable Diode Laser Diagnostic System for Collaborative Research on Air-Breathing Combustion				5a. CONTRACT NUMBER	
				5b. GRANT NUMBER F49620-02-1-0286	
				5c. PROGRAM ELEMENT NUMBER 61103D	
6. AUTHOR(S) Ronald K. Hanson				5d. PROJECT NUMBER 3484	
				5e. TASK NUMBER US	
				5f. WORK UNIT NUMBER	
7. PERFORMING ORGANIZATION NAME(S) AND ADDRESS(ES) Department of Mechanical Engineering Stanford University Stanford CA 94305-3030				8. PERFORMING ORGANIZATION REPORT NUMBER	
9. SPONSORING / MONITORING AGENCY NAME(S) AND ADDRESS(ES) AFOSR/NA 4015 Wilson Boulevard Room 713 Arlington VA 22203-1954				10. SPONSOR/MONITOR'S ACRONYM(S)	
				11. SPONSOR/MONITOR'S REPORT NUMBER(S)	
12. DISTRIBUTION / AVAILABILITY STATEMENT Approved for public release; distribution is unlimited					
13. SUPPLEMENTARY NOTES					
14. ABSTRACT This equipment grant focused on four areas: (1) portable diode laser sensors with new fiber-coupled diode lasers and the support equipment to provide higher power with extended wavelength tuning range and speed; (2) agile wavelength tuning to cover a broader wavelength region; (3) new wavelength sources for the ultraviolet to expand the wavelengths available for sensor strategy research; and (4) portable data acquisition and laser control electronics.					
15. SUBJECT TERMS diode laser spectroscopy, combustion					
16. SECURITY CLASSIFICATION OF:			17. LIMITATION OF ABSTRACT	18. NUMBER OF PAGES	19a. NAME OF RESPONSIBLE PERSON
a. REPORT Unclassified	b. ABSTRACT Unclassified	c. THIS PAGE Unclassified			Julian M. Tishkoff
			UL	5	19b. TELEPHONE NUMBER (include area code) (703) 696-8478

20030822 141

RECEIVED JUL 09 2003

Final Technical Report

**Portable Diode Laser Diagnostic System for
Collaborative Research on
Air-breathing Combustion**

AFOSR Grant F49620-02-1-0286

Prepared for

AIR FORCE OFFICE OF SCIENTIFIC RESEARCH

For the Period

June 1, 2002 to May 31, 2002

July 1, 2003

Submitted by

R. K. Hanson, Principal Investigator

Department of Mechanical Engineering
Stanford University
Bldg 530, Room 112
Stanford, CA 94305-3030

**HIGH TEMPERATURE GASDYNAMICS LABORATORY
Mechanical Engineering Department
Stanford University**

Portable Diode Laser Diagnostic System for Collaborative Research on Air-breathing Combustion

Ronald Hanson
Department of Mechanical Engineering
Stanford University, Stanford, CA 94305

EQUIPMENT ACQUIRED

Table 1 lists the equipment purchased by the DURIP Grant, Portable Diode Laser Diagnostic System for Collaborative Research on Air-breathing Combustion, AFOSR Grant F49620-02-1-0286 and matching funds from Stanford University Cost Sharing Accounts. This contract covered the period June 1, 2002 to May 31, 2002.

RESEARCH SUMMARY

Laser-based diagnostics measurements have enabled extraordinary advances in our understanding of the reactive flows during the past two decades. This understanding has led to the design and testing of new propulsion and aero-engine concepts. Modern propulsion test facilities are now commonly equipped with laser-based tools for flow visualization, spray characterization, and planar laser-induced fluorescence measurements of temperature, fuel mixing, and combustion intermediate species.

The development of inexpensive, rugged, solid-state laser sources by the telecommunications industry has led to a new revolution in practical and portable sensors for high-speed reacting flows. Simultaneous multi-parameter measurements using diode laser-based sensors now offer prospects for real-time combustion control. This equipment grant has provided equipment enabling research on portable sensors based on multiplexed telecommunications diode lasers. In addition, this grant provided equipment to extend the available wavelengths in the infrared and ultraviolet.

This equipment grant was focused on four areas: 1) portable diode laser sensors with new fiber-coupled diode lasers and the support equipment to provide higher power with extended wavelength tuning range and speed, 2) agile wavelength tuning to cover a broader wavelength region, 3) new wavelength sources for the ultraviolet to expand the wavelengths available for sensor strategy research, and 4) portable data acquisition and laser control electronics. The equipment list in Table 1 shows the distribution of expenses in these three focus areas.

This new equipment has allowed important expansion of our current studies. In particular, the portable diode laser sensor system assembled using this new equipment was used for experiments at Wright-Patterson AFB. Figure 1 illustrates the sensor used in the gas turbine combustor sector test rig to measure gas temperature using water vapor

absorption. Figure 2 shows the first-ever diode laser absorption measurements inside a gas turbine combustor sector; in this case of water vapor concentration and temperature.




Figure 1. Multiplexed diode laser sensor used to measure water vapor absorption in a gas turbine combustor sector test rig at WPAFB.




Figure 2. Temperature measured by the diode laser sensor varies with position with respect to the fuel injectors in the gas turbine combustor sector.

A similar experimental arrangement was also successful to measure time-resolved temperature in the scramjet combustor test rig at Wright-Patterson AFB. The collaboration with the scramjet team is continuing, and we anticipate additional measurement campaigns in the future.

We recently developed a new approach to UV absorption-based temperature sensing based on our recent discovery of strong temperature dependence in the CO₂ absorption cross-section. Initial tests of this strategy in the Stanford PDE are extremely encouraging; these measurements were enabled by the new high-power VUV lamp purchased on this equipment grant. Most recently the new cw 266 nm laser source funded by this grant was used to investigate absorption in shock-heated flows. The initial measurements are also quite encouraging; this source may prove a robust tool for routine absorption measurements in combustion flows.

Finally, the fabrication of a new spectroscopic data validation facility consisting of a new three-zone oven and test cells with gas mixture and pressure control is nearing completion. This facility will enable us to investigate the high temperature and elevated pressure spectroscopy needed to develop practical optical sensors to combustion and propulsion flows and provide controlled test conditions.

TABLE 1: Portable Diode Laser Diagnostic System for Collaborative Research on Air-breathing Combustion (Fabrication)

<i>ExpenditureFabrication</i>	<i>Cost (\$)</i>
Portable IR Laser Sensors	51,700
Agile Wavelength Tuning	97,400
Expanded Wavelength Range-UV	57,500
Portable Data Acquisition	29,900
TOTAL	\$236,500
Stanford University Contribution (10%)	\$ 23,650
Total AFOSR funds	\$212,850